
Female condoms — Requirements and test methods

AMENDMENT 1

Préservatifs féminins — Exigences et méthodes d'essai

AMENDEMENT 1





COPYRIGHT PROTECTED DOCUMENT

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Licensed by SOURCE to TBS TEXTILE SECTION
Order # NUMBER/Downloaded: 2023-09-12
Single-user licence only, copying and networking prohibited.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 157, *Non-systemic contraceptives and STI barrier prophylactics*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Female condoms — Requirements and test methods

AMENDMENT 1

Clause 15.3.3, item f)

Replace the phrase:

“if the manufacturer makes more than one size”

with:

“nominal width and nominal length of the female condom”.

Annex G

Add the following new subclause as G.2.7 and renumber the existing G.2.7 as G.2.8.

G.2.7 Test method verification

If required, verification that the test method is working correctly shall be confirmed by making holes in a number of individual containers using a hypodermic syringe needle with a diameter in the range 0,25 mm to 0,35 mm. Closely inspect the needle and individual container to confirm that the needle has completely penetrated right through the packaging material into the individual container.

A hypodermic needle in specified size range can become blunt after making as few as two holes. If a needle fails to penetrate into the individual container it should be replaced. If there is a failure to detect a hole made in the individual container the test should be repeated using a new needle before concluding that the test is not working.

Annex K

Add the following new subclause as K.4 and renumber the existing K.4 as K.5.

K.4 Test method verification

If required, verification that the test method is working correctly shall be confirmed by making holes in a number of female condoms using a hypodermic syringe needle with a diameter in the range 0,25 mm to 0,35 mm. Holes should be made near the closed end, in mid body and near the open end of the test condoms. Closely inspect the needle and condom to confirm that the needle has completely penetrated right through the material and into the condom.

